


## Analysis of m-Xylylenediamine (MXDA)



### Method Parameters

<b>Column</b>	Primesep AP, 4.6x150 mm, 5 µm, 100 Å
<b>Mobile Phase</b>	MeCN/H <sub>2</sub> O
<b>Buffer</b>	H <sub>2</sub> SO <sub>4</sub> – 0.2%
<b>Flow Rate</b>	1.0 mL/min
<b>Detection</b>	UV, 215 nm

Quelle: <https://sielc.com/Application-Analysis-of-m-Xylylenediamine-MXDA>