


## HPLC Analysis of Diphenamid on Primesep 100 and Obelisc R



Diphenamid is an herbicide which selectively controls weedy grasses and broadleaf plants in the production of tobacco, tomatoes, peppers, cotton, and many more commercial crops. Diphenamid is a preemergence herbicide that is surface applied to soils and can be expected to control weeds for 6-8 months. Diphenamid was analyzed on two different stationary phases. Primesep 100 is a reverse phase column that contains embedded acidic ion-pairing groups, while Obelisc R retains with long hydrophobic chains and multiple ionic pairing groups on the surface. Method is LC/MS compatible and can be replicated to analyze dozens of other pesticides.

SIELC has developed the Obelisc™ columns, which are mixed-mode and utilize Liquid Separation Cell technology (LiSC™). These cost-effective columns are the first of their kind to be commercially available and can replace multiple HPLC columns, including reversed-phase (RP), AQ-type reversed-phase, polar-embedded group RP columns, normal-phase, cation-exchange, anion-exchange, ion-exclusion, and HILIC (Hydrophilic Interaction Liquid Chromatography) columns. By controlling just three orthogonal method parameters - buffer concentration, buffer pH, and organic modifier concentration - users can adjust the column properties with pinpoint precision to separate complex mixtures.

## Method Parameters

<b>Column</b>	Primesep 100, 2.1×150 mm, 5 µm, 100 Å
<b>Mobile Phase</b>	Gradient MeCN – 10-70%
<b>Buffer</b>	Gradient AmAc pH 3.0- 20-60 mM
<b>Flow Rate</b>	0.4 mL/min
<b>Detection</b>	UV, 270 nm

Quelle: <https://sielc.com/Application-HPLC-Analysis-of-Diphenamid-on-Primesep-100-and-Obelisc-R>