


## HPLC Determination of Caffeine, Theophylline and Theobromine in Tea on Primesep SB Column



### Method Parameters

<b>Column</b>	Primesep SB, 3.2 x 100 mm, 5 µm, 100 Å, dual ended
<b>Mobile Phase</b>	EtOH – 10%
<b>Buffer</b>	H <sub>3</sub> PO <sub>4</sub> – 0.3%
<b>Flow Rate</b>	0.5 mL/min
<b>Detection</b>	UV, 270 nm

Quelle: <https://sielc.com/hplc-determination-of-caffeine-theophylline-and-theobromine-in-tea-on-primesep-sb-column>