


HPLC Method for Analysis of 1-Bromopropane on Newcrom R1 Column



Column:	Newcrom R1
Column size:	4.6 × 150 mm, 3µm 100A
Mobile phase:	MeCN/H ₂ O – 55/45%
Buffer:	None
Column Temp:	30°C
Flow rate:	1.0 mL/min
Detection:	UV 210 nm

Method Parameters

Column	Newcrom R1, 4.6×150 mm, 3 µm, 100 Å
Mobile Phase	MeCN – 55%
Buffer	None
Flow Rate	1.0 mL/min
Detection	UV 210 nm

Quelle: <https://sielc.com/hplc-method-for-analysis-of-1-bromopropane>