# HPLC Method for Analysis of 8-hydroxyquinoline with Primesep 100




## Method Parameters

| | |
|---|---|
| **Column** | Primesep 100, 4.6 x 150 mm, 5 µm, 100 Å, dual ended |
| **Mobile Phase** | MeCN – 20% |
| **Buffer** | H2SO4 – 0.3% |
| **Flow Rate** | 1.0ml/min |
| **Detection** | UV, 250 nm |

Quelle: https://sielc.com/hplc-method-for-analysis-of-8-hydroxyquinoline

AugustaChrom - Dr. Roland Kießwetter
Tel.: +49 821 599 783 70
Fax: +49 821 599 783 71
E-Mail: info@hplc-shop.de