## HPLC Method for Analysis of Allura Red in Soft Drinks 3




SIELC has developed the Obelisc™ columns, which are mixed-mode and utilize Liquid Separation Cell technology (LiSC™). These cost-effective columns are the first of their kind to be commercially available and can replace multiple HPLC columns, including reversed-phase (RP), AQ-type reversed-phase, polar-embedded group RP columns, normal-phase, cation-exchange, anion-exchange, ion-exclusion, and HILIC (Hydrophilic Interaction Liquid Chromatography) columns. By controlling just three orthogonal method parameters - buffer concentration, buffer pH, and organic modifier concentration - users can adjust the column properties with pinpoint precision to separate complex mixtures.

### Method Parameters

| | |
|---|---|
| **Column** | Obelisc N, 3.2×50 mm, 5 µm, 100 Å |
| **Mobile Phase** | MeCN/H2O – 50/50% |
| **Buffer** | AmFm pH 3.0 – 70 mM |
| **Flow Rate** | 0.6 mL/min |
| **Detection** | UV, 525 nm |

Quelle: https://sielc.com/hplc-method-for-analysis-of-allura-red-in-soft-drinks-3