

Method Parameters

Column	Primesep 100, 4.6 x 150 mm, 5 µm, 100 Å, dual ended
Mobile Phase	MeCN – 70%
Buffer	H2SO4 – 0.1%
Flow Rate	1.0 mL/min
Detection	Vis 408 nm
Injection Volume	5 µl

Quelle: <https://sielc.com/hplc-method-for-analysis-amphotericin-b>