


## HPLC Method for Analysis of Bis(2,2,6,6-tetramethyl-4-piperidyl) sebacate on Primesep B Column



<b>Column:</b>	Primesep B
<b>Column size:</b>	4.6 × 150 mm, 5 µm, 100 Å
<b>Column part number:</b>	B-46.150.0510
<b>Mobile phase:</b>	Gradient MeCN 10-70%, 10 min
<b>Buffer:</b>	H <sub>2</sub> SO <sub>4</sub> - 0.2%
<b>Flow rate:</b>	1.0 mL/min
<b>Detection:</b>	UV 210 nm

### Method Parameters

<b>Column</b>	Primesep B, 4.6×150 mm, 5 µm, 100 Å
<b>Mobile Phase</b>	Gradient MeCN
<b>Buffer</b>	H <sub>2</sub> SO <sub>4</sub> – 0.2%
<b>Flow Rate</b>	1.0 mL/min
<b>Detection</b>	UV 210 nm

Quelle: <https://sielc.com/hplc-method-for-analysis-of-bis2266-tetramethyl-4-piperidyl-sebacate-on-primesep-b-column>