


HPLC Method for Analysis of Bromhexine on Primesep 200 Column



Column: Primesep 200
Column size: 4.6 × 150 mm, 5 µm
Column part number: 200-46.150.0510
Mobile phase: MeCN/H₂O – 60/40%
Buffer: H₂SO₄ - 0.05%
Flow rate: 1.0 mL/min
Detection: UV 210, 245, 315 nm

Separation type: Liquid Chromatography Mixed-mode

Method Parameters

Column	Primesep 200 , 4.6×150 mm, 5 µm, 100 Å
Mobile Phase	MeCN/H ₂ O – 60/40%
Buffer	H ₃ PO ₄ – 0.05%
Flow Rate	1.0 mL/min
Detection	UV 210, 245, 315 nm

Quelle: <https://sielc.com/hplc-determination-of-bromhexine>