


## HPLC Method for Analysis of Propofol on Newcrom R1 Column



### Method Parameters

<b>Column</b>	Newcrom R1, 3.2 x 150 mm, 3 µm, 100 Å, dual ended
<b>Mobile Phase</b>	MeCN – 70%
<b>Buffer</b>	Phosphoric Acid
<b>Flow Rate</b>	0.5 mL/min
<b>Detection</b>	UV 270 nm

Quelle: <https://sielc.com/separation-of-propofol-on-newcrom>