


HPLC of Determination of 1,2,4-Triazole on Primesep 100 Column



Method Parameters

Column	Primesep 100, 4.6 x 150 mm, 5 µm, 100 Å, dual ended
Mobile Phase	MeCN/H ₂ O – 20/80%
Buffer	H ₃ PO ₄ – 0.1%
Flow Rate	1.0 mL/min
Detection	UV 200 nm

Quelle: <https://sielc.com/hplc-of-determination-124-triazole>