


## HPLC Separation of Isopropylamine and Glyphosate



### Method Parameters

<b>Column</b>	Primesep B, 4.6x50 mm, 5 µm, 100 Å
<b>Mobile Phase</b>	MeCN/H <sub>2</sub> O
<b>Buffer</b>	TFA
<b>Flow Rate</b>	1.0 mL/min
<b>Detection</b>	ELSD

Quelle: <https://sielc.com/Application-HPLC-Separation-of-Isopropylamine-and-Glyphosate>