


HPLC Separation of Neurotransmitters



The neurotransmitters dl-DOPA, creatinine, hydroxytryptophan (5-HTP), and serotonin are separated in less than 4 minutes on a short 50 mm Primesep 200 column. The HPLC separation uses a mobile phase of water, acetonitrile (MeCN, ACN) and trifluoroacetic acid (TFA) and UV detection at 250 nm. Ion-pair reagents were not needed for retention of these polar, hydrophilic compounds, instead, a combination of ion-exchange and reversed-phase interactions were used.

Method Parameters

Column	Primesep 200, 4.6x50 mm, 5 μ m, 100 Å
Mobile Phase	MeCN/H ₂ O – 80/20%
Buffer	TFA – 0.05%
Flow Rate	1.0 mL/min
Detection	UV, 250 nm

Quelle: <https://sielc.com/Application-HPLC-Separation-%20of-Neurotransmitters>