


MS-compatible HPLC Separation of Diphosphate Anion

Column: Primesep SB
Size: 4.6 x 50 mm
Mobile phase: MeCN -
10%, gradient of AmFm pH 3.0
from 20 to 100 mM in 7 min
Flow: 1.0 mL/min
Detection: ELSD 80°C



Method Parameters

Detection	ELSD Detection
-----------	----------------

Quelle: <https://sielc.com/Application-MS-compatible-HPLC-Separation-of-Diphosphate-Anion>