


Separation of Isomers of Aminobenzoic Acid



Method Parameters

Column	Primesep 100, 4.6×150 mm, 5 μm, 100 Å
Mobile Phase	MeCN
Buffer	H ₃ PO ₄ , H ₂ SO ₄
Flow Rate	1.0 mL/min
Detection	UV, 230 nm

Quelle: <https://sielc.com/Application-Separation-of-Isomers-of-Aminobenzoic-Acid>