


HPLC Method for Analysis of Lipitor

<https://sielc.com/HPLC-Method-Analysis-Lipitor>

Chromatogram



Method Parameters

Mobile Phase	Gradient MeCN – 30-65%, 5 min, 4 min hold
Buffer	Gradient H3PO4 0.05 – 0.3% , 5 min, 4 min hold
Flow Rate	1.0 ml/min
Detection	UV, 270 nm
Class of Compounds	Drug, Statin, Hydrophobic, Ionizable
Analyzing Compounds	Atorvastatin, Lipitor

HPLC Column Used

Primesep 200, 3.2×100 mm, 5 µm, 100A

[Order this column at hplc-shop.de →](#)