


HPLC Separation of Isopropylamine and Glyphosate

<https://sielc.com/Application-HPLC-Separation-of-Isopropylamine-and-Glyphosate>

Chromatogram



Method Parameters

Mobile Phase	MeCN/H ₂ O
Buffer	TFA
Flow Rate	1.0 ml/min
Detection	ELSD
Class of Compounds	Insecticide, Herbicide, Fungicide, Hydrophobic, Ionizable
Analyzing Compounds	Glyphosate, Isppropylamine

HPLC Column Used

Primesep B, 4.6x50 mm, 5 µm, 100A

[Order this column at hplc-shop.de →](http://hplc-shop.de)