


HPLC Separation of Parabens and Benzoic Acid

<https://sielc.com/Application-HPLC-Separation-of-Parabens-and-Benzoic-Acid>

Chromatogram



Method Parameters

Mobile Phase	MeCN/H ₂ O
Flow Rate	1.0 ml/min
Detection	UV, 210 nm
Buffer	H ₃ PO ₄
Class of Compounds	Acid, Hydrophilic, Preservative
Analyzing Compounds	Methyl paraben, Benzoic Acid, Propyl paraben

HPLC Column Used

Primesep B2, 4.6×150 mm, 5 µm, 100A

[Order this column at hplc-shop.de →](http://hplc-shop.de)